

20页



扫微信二维码
关注谱尼测试



160000343608

检 测 报 告

No. A2A601051006Z

委托单位 北京南宫生物质能源有限公司

受测单位 北京南宫生物质能源有限公司

签发日期 2020年07月11日

声 明
Statement

1. 本报告无检验检测专用章、报告骑缝章和批准人签章无效。
This report is invalid without special seal of inspection, cross-page seal and the approver's signatures.
2. 本报告页面所使用“PONY”、“谱尼”字样为本单位的注册商标，其受《中华人民共和国商标法》保护，任何未经本单位授权的擅自使用和仿冒、伪造、变造“PONY”、“谱尼”商标均为违法侵权行为，本单位将依法追究其法律责任。
The pattern and characters of "PONY" and "谱尼" used in this report are protected by the trademark law of the People's Republic of China. Any unauthorized usage, counterfeit, forgery and alteration of trademarks of "PONY" and "谱尼" are the violations of the law. The PONY has the right to pursue all legal liabilities of the subject of the delict.
3. 委托单位对报告数据如有异议，请于报告完成之日起十五日内（初级农产品报告请于报告收到之日起五日内）向本单位书面提出复测申请，同时附上报告原件并预付复测费。
If the applicant has any questions about the results, shall provide a written retest application with the original report, and prepay the retest fees to PONY within fifteen days since the approval date (as an exception, it shall be within five days since the date received fbr the primary agriculture products report).
4. 委托单位办理完毕以上手续后，本单位会尽快安排复测。如果复测结果与异议内容相符，本单位将退还委托单位的复测费。
After the applicant finishes the procedure mentioned above, PONY shall arrange the retest as soon as possible. If the retest result accords with the applicant dissent, PONY shall refund the retest fees.
5. 不可重复性或不能进行复测的实验，不进行复测，委托单位放弃异议权利。
Tests that can not be repeated and tested shall not be carried out again.
6. 委托单位对样品的代表性和资料的真实性负责，否则本单位不承担任何相关责任。
The applicant should undertake the responsibility for the provided samples' representativeness and document authenticity. Otherwise, PONY has not any relevant responsibilities.
7. 本报告仅对所测样品负责，报告数据仅反映对所测样品的评价，对于报告及所载内容的使用、使用所产生的直接或间接损失及一切法律后果，本单位不承担任何经济和法律责任。
This report is only responsible for the provided sample. The test results only represent the evaluation of the tested sample. PONY will not be responsible for any economical or legal liability generated from direct or indirect usage of the test report.
8. 本单位有权在完成报告后按规定方式处理所测样品。
PONY has the right to dispose the tested sample by rules, after approval of the test report.
9. 本单位保证工作的客观公正性，对委托单位的商业信息、技术文件等商业秘密履行保密义务。
PONY assures objectivity and impartiality of the test, and fulfills the obligation of confidentiality fbr applicants commercial information, and technique document.
10. 本报告私自转让、盗用、冒用、涂改、未经本单位批准的复制（全文复制除外）或以其它任何形式的篡改均属无效，本单位将对上述行为严究其相应的法律责任。
The report is invalid in case of illegal transfer, embezzlement, imposture, modification or any altering, reproducing except in full, without approval of PONY PONY shall investigate and affix the applicants legal liability accordingly.

▲防伪说明(Anti-counterfeiting Description):

- (1) 报告编号是唯一的；
The test report has exclusive report code.
 - (2) 报告采用特制防伪纸张印制，纸张表面带有“PONY”防伪纹路，该防伪纹路不支持复印，即复印件不会带有“PONY”防伪纹路。
The test report is printed by anti-copying paper whose surface shows "PONY" security print with specific anticounterfeiting technique. Security print will disappear after copying. Duplicates are not expected to give "PONY" security print under any circumstances.
- *****



全国服务热线
400-819-5688

WWW.PONYTEST.COM

北京实验室: (010) 83055000

武汉实验室: (027) 83997127

哈尔滨实验室: (0451) 58627755

上海实验室: (021) 64851999

长春实验室: (0431) 85150908

石家庄实验室: (0311) 85376660

温州实验室: (0577) 88271060

青岛实验室: (0532) 88706866

大连实验室: (0411) 87336618

乌鲁木齐实验室: (0991) 6684186

合肥实验室: (0551) 63843474

深圳实验室: (0755) 26050909

郑州实验室: (0371) 69350670

呼和浩特实验室: (0471) 3450025

广州实验室: (020) 89224310

天津实验室: (022) 23607888

西安实验室: (029) 89608785

杭州实验室: (0571) 85806807

厦门实验室: (0592) 5568048

苏州实验室: (0512) 62997900


太原实验室: (0351) 7555762

宁波实验室: (0574) 87977185

扫描二维码

关注谱尼测试微信

公众号 PONY4008195688



检测报告

No. A2A601051006Z

第 1 页, 共 1 页

委托单位	北京南宫生物质能源有限公司		
受测单位	北京南宫生物质能源有限公司		
受测地址	北京市大兴区青云店镇南大红门村南宫生活垃圾焚烧厂		
采样日期	2020-06-04	检测日期	2020-06-04~2020-06-10
排气筒名称	2#飞灰废气排气筒	标态干废气流量(m ³ /h)	905
排气筒高度(m)	36	废气平均温度(℃)	33
大气压(kPa)	100.3	废气平均流速(m/s)	3.8
采样位置	净化后	净化器厂家/名称/型号	——
样品编号	A2A02238-01	净化方式	布袋除尘
检测方法	固定污染源排气中颗粒物测定与气态污染物采样方法 GB/T 16157-1996 固定污染源废气 低浓度颗粒物的测定 重量法 HJ 836-2017		
检测仪器	见附表		
检测项目	排放浓度结果(mg/m ³)	排放速率结果(kg/h)	
颗粒物	1.9	1.72×10 ⁻³	

附表:

检测仪器 (名称、型号、公司编号)

设备名称	设备型号	公司编号
低浓度自动烟尘烟尘测试仪	ZR-3260D	IE-4482
恒温恒湿称重系统	8061B	IE-4281
电子天平	MS105DU	IE-2074

备注: 该报告中检测方法由委托单位指定。

——以下空白——

编制:

高钰彤

审核:

任阳

批准:

李玉强

